

in dispensing prescription drugs, and bearing labels containing the words "Professional Sample," or similar wording, and the names and addresses of manufacturers, packers, or distributors located outside the State of Pennsylvania. The bottle labeled *Deronil* had been repacked by the dealer.

**LIBELED:** 8-18-61, W. Dist. Pa.

**CHARGE:** 502(a)—while held for sale, the words "Professional Sample," and similar wording on the labels of a number of articles, were false and misleading as applied to these articles then in possession of a repacker and intended for sale, and not then intended for use as "complimentary -- not for sale" samples for physicians and others lawfully engaged in dispensing prescription drugs; and 502(f)(1)—the labeling of the article of drug labeled *Deronil* failed to bear adequate directions for use and it was not exempt from that requirement since it was a drug subject to the provisions of 503(b)(1) and its label failed to bear the correct identifying lot number as required by regulations.

**DISPOSITION:** 10-13-61. Default—destruction.

**6748. Electronic Magnetic Model G device.** (F.D.C. No. 46061. S. Nos. 26-992/93 R.)

**QUANTITY:** 3 devices at Ontario, Calif.

**SHIPPED:** 4-11-50 and 4-11-53, from Tiffin, Ohio.

**RESULTS OF INVESTIGATION:** The article was a suitcase-type unit which on opening displayed a control panel and detector plates. The control panel contained an array of switches, dials, push buttons, electrode terminals, and indicator lights. The electronic components within the case formed a power supply, oscillator, and amplifier for the detection and/or operation of hertzian waves.

**LIBELED:** 7-7-61, S. Dist. Calif.

**CHARGE:** 502(f)(1)—while held for sale, the labeling of the article failed to bear adequate directions for use in the diagnosis or treatment of any disease conditions, and it was not feasible to devise any directions for use because the article was worthless for any medical purposes.

**DISPOSITION:** 8-16-61. Default—delivered to the Food and Drug Administration.

**6749. Electronic Magnetic Model G device.** (F.D.C. No. 46065. S. No. 25-436 R.)

**QUANTITY:** 1 *Electronic Magnetic Model G device* at Yucaipa, Calif.

**SHIPPED:** 6-27-60, from Tiffin, Ohio, by L. L. Roby Manufacturing Corp.

**LABEL IN PART:** (Front) "Electronic Magnetic Model G" and (back) "Manufactured by L. L. Roby Manufacturing Corp. Tiffin, Ohio."

**ACCOMPANYING LABELING:** One instruction leaflet entitled "Electronic Magnetic Instrument Model G."

**RESULTS OF INVESTIGATION:** The article was a suitcase-type unit which on opening displayed a control panel and detector plates. The control panel contained an array of switches, dials, push buttons, electrode terminals, and indicator lights. The electronic components within the case formed a power supply, oscillator, and amplifier for the detection and/or operation of hertzian waves.

**LIBELED:** 7-7-61, S. Dist. Calif.